

ATTESTATION OF CONFORMITY

Certificate No: MDD-168

In conformance to the European Economic Commission 93/42/EEC Medical Devices Directive on harmonisation of laws, regulations and administrative documentation of Member States on Medical Devices and European Economic Commission directive 93/68/EEC amending Medical Devices Directive dated 22 July 1993,

the products manufactured by

İBRAHİM YILDIZ TEKSTİL LTD. ŞTİ.

at the following address

Eskihisar Mah. Ankara Bulvarı No:313 Z-1 Merkezefendi Denizli/TÜRKİYE

EN 13795-1:2019 Surgical Clothing and Drapes - Requirements and Test Methods - Part 1: Surgical Drapes and Gowns

Brand Name: VIROUT

Model: VR-05

(Standard Performance) are tested according to the following initial type tests by the manufacturer

For the assessment of conformity, the following documents were also reviewed:

Laboratory test results for Microbial Penetration, Bioburden,
Bursting and Tensile Strengths (wet/dry)

UNIVERSAL CERTIFICATION has evaluated production, design, intended use, risk evaluation according to safety purpose, product itself and add-on components (if exists) and product technical drawings of the surgical gowns manufactured and designed for use to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures. With this certificate, it is approved that the product fulfils all essential requirements and the related rules of 93/42/EEC Medical Devices Directive (MDD) Class I are applied. The information on the packaging for the above listed products covers the necessary information stated in Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745. This information includes; performance level and other relevant information given in EN ISO 15223-1:2016 and EN 1041:2008+A1:2013. It is considered to be suitable to attach a CE mark, as seen below, on your products in accordance with the information given in this certificate with publishing an EU Declaration of Conformity.

This certificate is issued on 01/07/2020 and valid until 30/06/2021 with the conditions that no change has been made with the product references and no change in the production process or not suspended or withdrawn for any reason.

ISTANBUL – 01/07/2020



Suat KAÇMAZ
UNIVERSAL CERTIFICATION
Director



Verify the validity with the QR Code

EU DECLARATION OF CONFORMITY

MANUFACTURER

İBRAHİM YILDIZ TEKSTİL LTD. ŞTİ.

Eskihisar Mah. Ankara Bulvarı No:313 Z-1 Merkezefendi Denizli/ TURKEY

PRODUCT DESCRIPTION

Brand Name: VIROUT

Model: VR-05

Surgical Gowns with standard performance to be used to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures, classified as Medical Device (Class I)

The Manufacturer declares on his sole responsibility that the product above is, under conditions of normal use and conditions defined by the Producer / the Manufacturer, safe and meets all the necessary legal conditions and requirements. The product, a medical device that is intended for single use and solely in accordance with the Manufacturer's instructions.

The Conformity is assessed especially with the following provisions:

- European Regulation (EU) 2017/745 and 93/42/EEC Medical Devices Directive establishing technical requirements for medical devices, in effective wording
- Technical standard EN 13795-1:2019 Surgical clothing and drapes - Requirements and test methods - Part 1: Surgical drapes and gowns
- Other relevant harmonized legislation and standards
- For the assessment of conformity, the following documents were also applied to:
- Results of laboratory tests for Microbial Penetration - Wet and Microbial Cleanliness, Bioburden by Ekoteks Laboratuvar ve Gözetim Hizmetleri A.Ş.
- Results of laboratory tests for Bursting and Tensile Strengths (wet/dry) by Ekoteks Laboratuvar ve Gözetim Hizmetleri A.Ş.

MARKING, LABELLING

Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745 specifies the information that should be specified on the packaging in which the surgical gown is supplied. The information supplied with the product considering EN ISO 15223-1:2016 and EN 1041:2008+A1:2013.


MEASURES TO ENSURE CONFORMITY

The Manufacturer declares that he has taken all necessary measures to ensure the conformity of products placed on the market with technical documentation and basic requirements for this type of product.

İbrahim YILDIZ

General Manager

01/07/2020





**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE

TEST REPORT
DENEY RAPORU

20015490-
ing-Add

06-20

Customer name: İBRAHİM YILDIZ TEKSTİL SAN. TİC. LTD. ŞTİ.
Address: ESKİHİSAR MAH. ANKARA BULV. NO:313 Z-1 MERKEZEFENDİ
Buyer name: -
Contact Person: HALİM İNCEKOL
Order No: -
Article No: -
Name and identity of test item: White surgical gown
The date of receipt of test item: 15.05.2020
Re-submitted/re-confirmation date: -
Date of test: 15.05.2020-25.06.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 7



Date
25.06.2020

Customer Representative
Servin YURTSEVEN

Head of Testing Laboratory
Sevim A. RAZAK
25.06.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

20015490-
ing-Add

06-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TEST		
Microbial Cleanliness (Bioburden)	P	
Wet-Bacterial Penetration	P	
Dry-Bacterial Penetration ⁽¹⁾	P	
PHYSICAL PROPERTIES TESTS		
Tensile Strength / Dry	P	
Tensile Strength / Wet	P	
Bursting Strength / Dry	P	
Bursting Strength / Wet	P	
P: Pass F: Fail R: Refer to retailer technologist. Test results were evaluated according to EN 13795-1:2019 Standard Performance Properties Critical Sample Group limit values (Table 1) ⁽¹⁾ This report was reissued to add this test result		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



Gen.f136-2/03

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

TEST RESULTS

Surgical clothing and drapes - Requirements and test methods – Part 1: Surgical drapes and gowns EN 13795-1 :2019

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: EN ISO 11737-1:2018 /TS EN ISO 11737-1 :2018 (*)

The sample is put in extraciton liquid after shaking well after shaking well (250 rpm,5 min), inoculated on the suitable agar.The plates are incubated for 3 days at 30 ± 1 ° C for 72 hours, and 7 days at (20 to 25) °C for TSA and SDA plates respectively. Total microoragnisms counts are calculated.

	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	160 cfu/100 cm ²	≤300 cfu/100 cm ²

*cfu= Colony forming unit.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

20015490-
ing-Add

06-20

TEST RESULTS

WET-BACTERIAL PENETRATION

Test Method: BS EN 22610: 2006 (Surgical drapes, garments and fresh air clothes used as medical devices for patients, hospital staff and equipment - Test method for determination of resistance to wet bacterial permeability) (*)

A test sample is placed on the agar plate on a rotating disc. Bacteria carrier material and coating film are placed on the test sample and all parts are fixed on the disk. A finger is placed on the test sample to apply a certain force ($3N \pm 0.02$). The finger moves on the test sample over the entire surface of the agar within 15 minutes. 5 studies are carried out for 15 minutes. 6. The study is repeated by inverting the sample.

Sample amount:	5 pieces 25x25cm ²
Carrier Material:	30 µm thin, 25x25cm ² Polyurethane Film
Coating Material:	25x25cm ² HDPE Film
Microorganism:	Staphylococcus aureus ATCC 29213
Bacterial Concentration (kob / ml):	1-4x10 ⁴ kob / ml
Incubation Conditions:	(36 ± 1) ° C 48 hours

RESULTS			
Number of Populating Bacteria (cfu)		Penetration Rate	
X₁	0	R_{CUM1}	0
X₂	0	R_{CUM2}	0
X₃	0	R_{CUM3}	0
X₄	130	R_{CUM4}	0,25
X₅	154	R_{CUM5}	0,55
Z	230		
T		514	

X₁ X₅: Number of colonies growing in 5 parallel petri in the same sample

Z: number of colonies growing in the sixth petri dish

T: X₁ + X₂ + X₃ + X₄ + X₅ + Z

$$R_{CUM1} = X_1/T$$

$$R_{CUM2} = (X_2 + X_1)/T$$

$$R_{CUM3} = (X_3 + X_2 + X_1)/T$$

$$R_{CUM4} = (X_4 + X_3 + X_2 + X_1)/T$$

$$R_{CUM5} = (X_5 + X_4 + X_3 + X_2 + X_1)/T$$

BARRIER INDEX (I_B)

	Result	Expected value (*)
I_B	5,19	≥2,8

$$I_B = 6 - (CUM1 + CUM2 + CUM3 + CUM4 + CUM5)$$

* EN 13795-1:2019 Surgical gowns and drapes - Requirements and test methods are evaluated according to Table-1.

Gen.f136-2/03

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

20015490-
ing-Add

06-20

Test Method: ISO 22612: 2005 (Clothing for protection against infectious agents - Test method for resistance to dry microbial penetration)

Samples and containers are sterilized. Agar plates are placed in each container. Samples are placed aseptically in the apparatus. The covers are closed. After making a pot in the sample with the piston, the pistons are removed and 0.5 g ± 0.1 g are added to five samples from the powder contaminated with bacteria and the six to the non-contaminated powder. Then all openings are closed with a plastic bag. The device is operated to give 20,800 vibrations per minute. The test time is 30 minutes. After the test is over, all agar plates are incubated at 35 ° C for 24 hours.

Sample amount:	6 pieces 20x20 cm ²
Mikroorganism:	<i>Bacillus subtilis</i> ATCC 9372
Incubation conditions:	35°C / 24 hours
RESULTS	
Number of Populating Bacteria (cfu)	
1	1
2	5
3	12
4	9
5	18
6 (Control)	0
Total	45
Logarithm	1,65
EVALUATION	
Result	Class (*)
1 < log kob ≤ 2	2
<i>* EN 14126: 2003 Protective Clothing - Performance Properties and Test Methods of Protective Clothing Against Infectious Agents are evaluated according to Table-4.</i>	
Sınıf	Penetrasyon (log kob)
3	≤ 1
2	1 < log kob ≤ 2
1	2 < log kob ≤ 3
<i>* EN 13795-1:2019 Surgical gowns and drapes - Requirements and test methods are evaluated according to Table-1.</i>	
RESULT	
Result (cfu/g)	Expected Value
45 cfu/g	≤300 cfu/g

Gen.f136-2/03

20015490-
ing-Add

06-20

TEST RESULTS

TENSILE STRENGTH; EN 29073-3:1996 (*)

Instron 5969 (Load: 5 kN), Strip Method.
Speed: 100 mm/min±10, Gauge length 200 mm.
Pre-load was not applied. Without wetting samples.
The average results are given for width and length direction of four samples
Performed in the conditioned room (20±2°C-65%±4).

Dry ;

	<u>RESULT</u>
Width	52.7 N
Length	91.5 N

REQUIREMENT

≥ 20N (Dry)
≥ 20N (Dry)

TENSILE STRENGTH; EN 29073-3:1996 (*)

Instron 5969 (Load: 5 kN), Strip Method.
Speed: 100 mm/min±10, Gauge length 200 mm.
Pre-load was not applied. With wetting samples.
The average results are given for width and length direction of four samples
Performed in the conditioned room (20±2°C-65%±4).

Wet ;

	<u>RESULT</u>
Width	53.6 N
Length	94.9 N

REQUIREMENT

≥ 20N (Wet)
≥ 20N (Wet)

BURSTING STRENGTH;; ISO 13938-1:1999

SDL ATLAS M229 tester. Test area: 30.5 mm diameter
The average results are given of five samples.
Performed in the conditioned room (20±2°C-65%±4).

	<u>RESULT</u>
Dry ;	145.8 kPa
Height at Burst*	10.4 mm

REQUIREMENT

≥ 40 kPa (Dry)

20015490-
ing-Add

06-20

TEST RESULTS

BURSTING STRENGTH;; ISO 13938-1:1999

SDL ATLAS M229 tester. Test area: 30.5 mm diameter
The average results are given of five samples.
Performed in the conditioned room (20±2°C-65%±4).

	<u>RESULT</u>	<u>REQUIREMENT</u>
Wet ;	155.6 kPa	≥ 40 kPa (Wet)
Height at Burst*	11.0 mm	